

**PASTEUR INSTITUTE OF INDIA,**  
**COONOOR - 643 103, THE NILGIRIS.**  
(Ph.0423 2231852, Fax - 0423-2231655)

**CORRIGENDUM**

Dated: 06/03/2019

**CHANGES MADE IN URS (USER RECOMMENDED SPECIFICATION) AND EXTENSION  
OF DUE DATE FOR SUBMISSION OF QUOTATION IN RESPECT OF Tender  
Notification No.PII/ADVT/001/2019 dt.28-01-2019**

-----

As per our Tender Notification No.PII/ADVT/001/2019 dated 28/01/2019 and Corrigendum dated 22/02/2019 for the supply of the 11 Nos of Air Sampler, the URS mentioned in our tender has been amended as detailed below and the due date for submission of bid has been extended as detailed below: -

1.	<b><u>Specification of the material of construction:</u></b> Point No.2.1.1 is amended as - with water proof, corrosion resistant, easy to sanities, compatible and resistance to chemicals, powder coated with aluminium body or any other suitable material of construction. It shall be used in cleanroom.
2.	2.1.2 is removed as this is pertinent to Head/Sieve already mentioned in point Nos.2.1.12 and 2.1.14. In Point Nos. 2.1.12 and 2.1.14 the Head/Sieve may be SS 316L, 316 or 304.

<b>S.No.</b>	<b>Description</b>	<b>Schedule</b>
i.	Closing date & time for receipt of Tender	21-03-2019, 15.00 Hrs
ii.	Time and date of opening of Technical Bids	21-03-2019, 15:30 Hrs
iii.	Venue of opening of Tender	PASTEUR INSTITUTE OF INDIA, COONOOR.

Other Specifications, Terms and conditions for submission of bids mentioned in our above notification remain the same.

**Sd/-**  
**DIRECTOR**